

QA File Parameter

Patient Name : XXXXXXXXXX
 Patient ID : XXXXXXXXXX
 Plan Date : 2/5/2020
 SSD :
 SDD :
 Depth :
 Energy : 6X
 Angle :

Absolute Dose Comparison

Difference (%) : 2.0
 Distance (mm) : 2.0
 Threshold (%) : 20.0
 Meas Uncertainty : No
 Use Global % : No

Summary (GC Analysis)

Total Points : 2197
 Passed : 1940
 Failed : 257
 % Passed : 88.3
 *GC/Gamma is using 3D Mode

Dose Values in cGy


	CAX	Norm	Sel	Max
AC Merged		153.76	60.50	153.76
Plan		158.70	59.31	158.94
AC Merged		-4.94	1.19	-5.19
-Plan				
% Diff		-3.12	2.01	-3.26
Coords (y,x) cm	0,0	-9,-28.5	0,-0.5	

Notes


Fields 1-2
 Pelvis=Cervix+PM+LNs
 SLF

Reviewed By :

AC Merged



Plan



AC Merged-Plan

